



Instruction manual

TE67X000009637 DeviceNet Module

- For transmitting status and weight of digital load cells

Instruction manual no.: IM-TE91K014-EN2 ESE02424EN Date of issue: September 9, 2014

First published: March 8, 2013

Original manual

1) Contents

1) Contents	1
2) Introduction	3
2.1 Introduction	3
2.2 Target group	3
2.3 Limitation of liability	4
2.4 Intended use	4
3) Data exchange	5
3.1 DeviceNet communication using PPO	5
3.2 IEEE754 format	7
4) Zeroing and calibration	8
4.1 Zeroing procedure	8
4.2 Calibration procedure	8
5) Installation of system	
5.1 Checklist during installation	9
6) TE67X000009637 hardware description	10
6.1 DIP-switch settings	10
6.2 Jumpers	11
6.3 Light emitting diodes	12
6.4 EE-bus connector	
6.5 DeviceNet connector	14
7) Maintenance, Service / Repair & Warranty	15
7.1 Maintenance	
7.2 Service / Repair	
7.3 Warranty conditions	
8) How to contact Alfa Laval Tank Equipment	16

2) Introduction

2.1 Introduction

This document describes the use of the Alfa Laval Kolding A/S TE67X00009637 DeviceNet communication module, when equipped with program listed on the front page. With this program the TE67X000009637 communication module is capable of transmitting weight and status of up to 4 loadcells in a single telegram. Each loadcell is connected to the communications module through a TE67X000009610 loadcell interfacemodule.

It is possible to connect the TE67X000009637 communication module to DeviceNet, where it will act as a slave. It will then be possible from the DeviceNet master to read status and weight for each of the connected loadcells. Functions as zeroing, calibration and calculation of system weight **must** be implemented on the DeviceNet master.

Exchange of data between DeviceNet master and slave is performed as described in the following.

Note: The illustrations and specifications contained in this manual were effective at the date of printing. However, as continuous improvements are our policy, we reserve the right to alter or modify any unit specification on any product without prior notice or any obligation.

© Alfa Laval Corporate AB

This document and its contents is owned by Alfa Laval Corporate AB and protected by laws governing intellectual property and thereto related rights. It is the responsibility of the user of this document to comply with all applicable intellectual property laws. Without limiting any rights related to this document, no part of this document may be copied, reproduced or transmitted in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise), or for any purpose, without the expressed permission of Alfa Laval Corporate AB. Alfa Laval Corporate AB will enforce its rights related to this document to the fullest extent of the law, including the seeking of criminal prosecution.

- Technical modifications reserved -

2.2 Target group

 Warning:
 To avoid operator hazards and damages of the device, the following instructions have to be worked out by qualified technical personnel.

2) Introduction (continued)

2.3 Limitation of liability

By non-observance of the instruction manual, inappropriate use, modification or damage, no liability is assumed and warranty claims will be excluded.

2.4 Intended use

The Alva Laval TE67X000009637 DeviceNet communication module has, according to the type, been developed for Tank weighing applications. It is the operator's responsibility to check and verify the suitability of the device for the intended application. If any doubts remain, please contact our local Alfa Laval Company in order to ensure proper usage. Alfa Laval Kolding A/S is not liable for any incorrect selections and their effects.

- The technical data listed in the current data sheet are engaging and must be complied with. If the data sheet is not available, please order or download it from our homepage (http://www.alfalaval.com)

3) Data exchange

3.1 DeviceNet communication using PPO

DeviceNet communication with the TE67X000009637 communication module uses a so called 'parameterprocess data object' (PPO) consisting of 26 bytes. This telegram (object) is only used when transferring data from the slave to the master, since no data are transmitted from the master to the slave. The structure for this telegram is as follows:

LC LC LC		-		Lc		Lc							
Register Status(0)		Weight(0)		Statu	s(3)	Wei	ght(3	5)					
0	1	2	3	4	5	6	7	20	21	22	23	24	25

The byte order (MSB/LSB first?) for the individual parts of the telegram is determined by jumper JU3. Normally this jumper is set from the factory so that LSB comes first. In the following bit 0 will represent the least significant bit in a register.

LcRegister is two bytes that constitute a bit register for indication of connected loadcells detected during power on. Hence bit 0-3 will be ON, if the corresponding loadcell (address) was detected during power on.

LcStatus(X) is two bytes that constitute a register containing the actual status for loadcell X. The individual bits in the status register have the following function:

3) Data exchange (continued)

BIT-NO	FUNCTION
0	Invalid/missing 'sample' ID
	Bad connection between communicationsmodule and loadcell module. Not all telegrams
	from the communicationsmodule are received in the LC module.
1	Timeout – Loadcell not connected ?
	Check that the loadcell is connected to the loadcell module.
2	Loadcell not synchronised
	Bad connection between loadcell and loadcell module or very heavy over- or underload.
3	Hardware synchronisation error
	Loadcell samplings are not synchronised.
4	Power failure
	The supply voltage is to low.
5	Overflow in weight calculation
	Internal error in loadcell module.
6	Invalid/missing 'latch' ID
	Bad connection between communicationsmodule and loadcell module. Not all telegrams
	from the communicationsmodule are received in the loadcell module.
7	No answer from loadcellmodule
	No data are received from this loadcell module. This can be caused by the removal of the
	loadcell module, no power to the module or that the connection between loadcell module
	and communicationsmodule is broken.

LcWeight(X) is four bytes constituting a register containing the actual weight signal from loadcell **X** in gram. Depending on JU4 the weight will be in the following format:

JU4=OFF:	in "32 bit signed integer" format.
JU4=ON:	in "IEEE754" format.

Since only status and weight for the loadcells are transmitted in the telegram functions such as status handling, calculation of system weight, zeroing and calibration **must** be implemented on the DeviceNet master.

Calculation of system weight is done by addition of the weight registers for the connected loadcells. Note that the result is only valid if all status registers for the connected loadcells indicate **no** errors. It should also be noted that it is up to the master to ensure the usage of consistent loadcell data when calculating the system weight (the used data should come from the same telegram).

3) Data exchange (continued)

3.2 IEEE754 format

If by usage of JU4 representation of data on IEEE754 format is selected this is done as follows:

Byte1 (MSB)			Byte2		Byte3		Byte4 (LSB)	
bit7	bit6 bit0	bit7	bit6	bit0	bit7	bit0	bit7	bit0
S	2 ⁷ 2 ¹	2 ⁰	2 ⁻¹	2 ⁻⁷	2 ⁻⁸	2 ⁻¹⁵	2 ⁻¹⁶	2 ⁻²³
Sign	Exponent		Man	tissa	Man	ntissa	Mai	ntissa

Byte1 (MSB)			Byte2		Byte3		Byte4 (LSB)		
bit7	bit6	bit0	bit7	bit6	bit0	bit7	bit0	bit7	bit0
S	2 ⁷	2 ¹	2 ⁰	2 ⁻¹	2 ⁻⁷	2 ⁻⁸	2 ⁻¹⁵	2 ⁻¹⁶	2 ⁻²³
Sign	Ex	ponent	onent		tissa	Ма	ntissa	Mar	ntissa

Formula:

Value = $(-1)^{S} * 2^{(exponent-127)} * (I+Mantissa)$

Example:

Byte1	Byte2	Byte3	Byte4
0100 0000	1111 0000	0000 0000	0000 0000
Value = $(-1)^{0} * 2$	2 ⁽¹²⁹⁻¹²⁷⁾ * (1 + 2 ⁻¹ + 2	$(-2^{-2} + 2^{-3}) = 7.5$	

Please note that if MSB first has been selected using JU3 the byte with the "sign" will come first in the weight indications, and if LSB first has been selected the byte with the "sign" will come last in the weight indications.

4) Zeroing and calibration

4.1 Zeroing procedure

Zeroing of the system (all loadcells) should be performed as follows:

- 1) The weighing arrangement should be empty and clean.
- 2) The DeviceNet master reads and stores the actual weight values for the connected loadcells in corresponding zeroing registers.
- 3) After this the actual weight for loadcell **X** can be calculated as:

LcGross(X) = LcWeight(X) - LcZero(X)

and the system weight (uncalibrated) for the connected loadcells is calculated as:

SystemWeight = LcGross(0) + LcGross(1) +

4.2 Calibration procedure

Fine calibration of the system should be performed as follows:

- Check that the weighing arrangement is empty, and that the gross weight is zero. Zero if necessary.
- 2) Place a known load (calibrationweight) on the weighing arrangement.
- Calculate the calibration factor that should be multiplied on the system weight in order to achieve correct showing as:

Calibrationfactor = (Calibrationweight)/(Actual showing)

After this the determined calibration factor is used to calculate the calibrated weight as follows:

Sys.Weight(Calibrated) = Calibrationfactor * Sys.Weight(Uncalibrated)

If the determined calibration factor falls outside the interval 0.9 to 1.1 it is very likely that there is something wrong with the mechanical part of the system.

5) Installation of system

5.1 Checklist during installation

During installation of the system the following should be checked:

- 1) If necessary the DeviceNet master should be configured to communicate with the DeviceNet module (TE67X000009637) using the supplied EDS file.
- The loadcells are mounted mechanically and connected to the DeviceNet module (TE67X00009637) using their corresponding loadcell interface module (TE67X000009610). The loadcell addresses are set using the DIP-switches (Sw1.5-Sw1.8) on the TE67X000009610 modules, so that they forth running from address 0 (0-3).
- 3) The DeviceNet module (TE67X000009637) is connected to the DeviceNet, and possibly a termination is made at this DeviceNet slave.
- The baudrate of the DeviceNet module (TE67X00009637) is set using Sw2.1-Sw2.2 and its address is set using Sw2.3- Sw2.8. Power is applied and the DeviceNet communication is started.
- 5) Verify that the yellow LED (D1) on the DeviceNet module (TE67X000009637) is lit, and that the red LED (D2) is not lit. Verify that the TXBB LED on the DeviceNet module is lit and that the TXBB LED's on the loadcell modules (TE67X000009610) are also lit (can flash slightly). Verify that both the MS and NS LED's end up being lit green constantly.
- 6) Verify that the DeviceNet module (TE67X000009637) has found the correct loadcells (LcRegister), and that no loadcell errors are indicated (LcStatus(x)).
- 7) Verify that every loadcell gives a signal (**LcWeight(x)**) by placing a load directly above each loadcell one after the other (possibly with a known load).

The system is now installed and a a possible zero and fine calibration is made as described earlier. Finally verify that the weighing system(s) returns a value corresponding to a known actual load.

Note that in the above checklist no consideration has been made on which functions are implemented on the DeviceNet master.

6) TE67X000009637 hardware description

6.1 DIP-switch settings

The TE67X000009637 module is equipped with two DIP-switch blocks. DIP-switch block 1 has the following function:

SWITCH	FUNCTION
Sw1.1-Sw1.4	Reserved for future use

DIP-switch block 2 has the following function:

SWITCH	FUNCTION
Sw2.1-Sw2.2	Setting of DeviceNet DataRate (DR)
	The desired baudrate is set according to the table shown below. Note that
	these switches are only read during power-on.
Sw2.3-Sw2.8	Setting of DeviceNet Node Address (NA)
	The address (0-63) is set as the DIP-switches are binary coded, so that Sw2.8
	is LSB and Sw2.3 is MSB. Note that these switches are only read during pow-
	er-on.

The baudrate of the TE67X000009637 module is set according to this table:

Sw2.2	Sw2.1	Baudrate
OFF	OFF	125 kbps
ON	OFF	250 kbps
OFF	ON	500 kbps
ON	ON	Not allowed

6.2 Jumpers

The TE67X000009637 module is equipped with 5 internal jumpers. These jumpers have the following function:

JUMPER	FUNCTION
JU2	Test mode
	JU2 OFF: Normal mode (Default at delivery. Should'nt be changed)
	JU2 ON: Test mode
	The jumper must be OFF during normal operation.
JU3	Selection of LSB/MSB data format
	The jumper determines the byte order in which data are transmitted/received.
	JU3 OFF: LSB first (normal setting from factory)
	JU3 ON: MSB first
JU4	Selection of (32 Bit Signed Integer) / (IEEE754) data format
	The jumper determines if the weight indications in the telegram are in 32 bit
	signed integer or in IEEE754 floating point format.
	JU4 OFF: 32 bit signed integer format (normal setting from factory)
	JU4 ON: IEEE754 floating point format
JU5	Reserved for future use
JU6	Test mode
	JU6 OFF: Normal mode (Default at delivery. Should'nt be changed)
	JU6 ON: Test mode
	The jumper <u>must</u> be OFF during normal operation.

6.3 Light emitting diodes

The TE67X000009637 module is equipped with 6 light emitting diodes (LED's). These LED's have the following function:

LED	FUNCTION
TxBB	Communication with loadcells
(Green LED)	The TE67X000009637 communicates with the loadcells
D1	DeviceNet Voltage Detected
(Yellow LED)	The TE67X000009637 module has detected DeviceNet voltage on the De-
	viceNet connector.
D2	Reserved for future use
(Red LED)	
TxCAN	CAN bus TxD (Transmit Data)
(Green LED)	The TE67X000009637 module transmits data across the CAN bus.
MS	Module Status LED
(Green/Red LED)	The TE67X000009637 Module Status LED, that can be lit/flashing in different
	colours depending on the status of the module. The function of the MS LED
	is given in the table below.
NS	Network Status LED
(Green/Red LED)	The TE67X000009637 Network Status LED, that can be lit/flashing in differ-
	ent colours depending on the status of the network. The function of the NS
	LED is given in the table below.

Please note that the LED's will flash shortly during power-up during the selftest of the module. The MS and NS LED's will shortly flash Green/Red. The MS and NS LED's can in conjunction with the table below be used for error finding.

Light emit-	Colour	Status	Description
ting diode			
MS	Green	ON	Normal Operation. Communication performed normally.
		Flashing	Standby State. The unit needs supervision.
	Red	ON	Unrecoverable fault. A timer error, memory error or other system
			error. The unit may need replacing.
		Flashing	Recoverable fault. Configuration error, DIP-switch not set correct
			or similar error. Correct error and restart unit.
		OFF	No power. The power is disconnected or the unit is being restart-
			ed.
NS	Green	ON	On-Line, Connection OK. The unit is On-Line and a connection
			with the master has been established.
		Flashing	On-Line, No Connection. The unit is On-Line but no connection to
			the master has been established.
	Red	ON	Critical Communication Error. The unit has detected an error that
			makes it impossible to communicate on the network (duplicate
			MAC Id or Bus-Off error).
		Flashing	Communication Time-Out. One or more I/O connections are in
			the Time-Out state.
		OFF	No power/Off-line. The device may not be powered.

6.4 EE-bus connector

The TE67X000009637 module is equipped with a 10 pole connector for connection to the Alfa Laval Kolding A/S EE-bus. Hereby connection to the individual TE67X000009610 loadcell modules as well as to the power supply for the TE67X000009637 module is achieved. The connection is made using a flat cable with mounted connectors for the individual modules. The 10 pole connector has the following connections:

<u>J4 Connector</u>	Function
J1.1-J1.2	RS485-B (negative line)
J1.3-J1.4	RS485-A (positive line)
J1.5-J1.6	0VDC (Gnd1)
J1.7-J1.8	+24VDC (Vin1)
J1.9-J1.10	I/O line

6.5 DeviceNet connector

The TE67X000009637 module is equipped with a 5 pole connector for connection to DeviceNet. The connection is according to the DeviceNet specification and is made as follows:

<u>J2 Connector</u>	Function	<u>Colour</u>
J2.1	V-	(Black)(0VDC input)
J2.2	CAN_L (Blue)	
J2.3	SHIELD	(Grey)
J2.4	CAN_H (White)	
J2.5	V+	(Red)(24VDC input)

7) Maintenance, Service / Repair & Warranty

7.1 Maintenance

This device is maintenance-free, but to ensure optimum accuracy of the weighing installation it is recommended to inspect the weighing system installation at regular intervals.

Recommended inspection points are:

1: Verify that tank is freestanding. Are all pipe connections flexible and not restraining vertical movement of the tank?

1a: Verify that no changes have been done to the tank? If any modifications has been performed ensure that they are done correctly according to installation recommendations for weighing systems.

2: Verify that all legs of the tanks all are in physical contact with load cells.

2a: on analog output module (TE67X000002029), verify that load from the empty tank is evenly distributed between the legs, load on each leg can be read-out through parameter "LCx mode". If load in unevenly distributed check and re adjust tank legs to ensure even load distribution.

7.2 Service / Repair

7.2.1 Return

Upon every return of the device, no matter if for recalibration, modifications or repair, it is necessary to contact your local Alfa Laval office to guarantee a quick execution of your request.

Please inform us by sending an email to:Alteq.PartsandService@alfalaval.com. Include the number of devices sent and request a Return Number. Afterwards clean the device, pack it shatterproof and send it to Alfa Laval Kolding A/S indicating the Return Number.

7.3 Warranty conditions

The warranty conditions are subject to the legal warranty period of 12 months from the date of delivery. In case of improper use, modifications of or damages to the device, we do not accept warranty claims. Damaged diaphragms will also not be accepted. Furthermore, defects due to normal wear are not subject to warranty services.

8) How to contact Alfa Laval Tank Equipment

For further information please feel free to contact:

Alfa Laval Tank Equipment

Alfa Laval Kolding A/S

31, Albuen - DK 6000 Kolding - Denmark
Registration number: 30938011
Tel switchboard: +45 79 32 22 00 - Fax switchboard: +45 79 32 25 80
www.toftejorg.com , www.alfalaval.dk - info.dk@alfalaval.com

Contact details for all countries are continually updated on our websites.

9) EC Declaration of Conformity

	EC Declaration of Conformity
	Eo Deciaration of comonnity
The designated company	
Alfa Laval Kolding A/S	
Albuen 31, 6000 Kolding, Den	ımark
Address + +45 79 32 22 00	
Phone no.	
hereby declare that	
Alfa Laval Device	eNet Module TE67X000009637
Denomina	
And furthermore declares that - EN 61010 Part 1 (Safety I - EN 60529 (degrees of pro	sions of the Directive 73/23/EEC appendix IIIB and directive 89/356/EEC. t the following (parts/clauses of) applicable directives have been used: requirements for electrical equipment for measurements control and laboratory use) otection provided by enclosures) eneric emission standard) pric immunity standard)
The technical construction file R&D Manager	e for this instrument is retained at the above address.
Title	Name Signature
Se	eptember 9, 2014 Alfa Laval Kolding A/S Date Company
	CE



© Alfa Laval Corporate AB

This document and its contents is owned by Alfa Laval Corporate AB and protected by laws governing intellectual property and thereto related rights. It is the responsibility of the user of this document to comply with all applicable intellectual property laws. Without limiting any rights related to this document, no part of this document may be copied, reproduced or transmitted in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise), or for any purpose, without the expressed permission of Alfa Laval Corporate AB. Alfa Laval Corporate AB will enforce its rights related to this document to the fullest extent of the law, including the seeking of criminal prosecution.